

八王子
2010年6月
検量線NO₂
館町0402

係数
0.0200

定数M3
6.9990E-05

定数M2
7.1235E-04

定数M1
0.20988821

電流値	NO ₂ 濃度	NO ₂ 濃度
μA	μg	ppm
0.0	0.00	0.000
0.5	0.11	0.007
1.0	0.21	0.009
1.5	0.32	0.011
2.0	0.42	0.013
2.5	0.53	0.015
3.0	0.64	0.016
3.5	0.75	0.017
4.0	0.86	0.019
4.5	0.97	0.020
5.0	1.08	0.021
5.5	1.19	0.022
6.0	1.30	0.023
6.5	1.41	0.024
7.0	1.53	0.025
7.5	1.64	0.026
8.0	1.76	0.027
8.5	1.88	0.027
9.0	2.00	0.028
9.5	2.12	0.029
10.0	2.24	0.030
10.5	2.36	0.031
11.0	2.49	0.032
11.5	2.61	0.032
12.0	2.74	0.033
12.5	2.87	0.034
13.0	3.00	0.035
13.5	3.14	0.035
14.0	3.27	0.036
14.5	3.41	0.037
15.0	3.54	0.038
15.5	3.69	0.038
16.0	3.83	0.039
16.5	3.97	0.040
17.0	4.12	0.041
17.5	4.27	0.041
18.0	4.42	0.042
18.5	4.57	0.043
19.0	4.73	0.043
19.5	4.88	0.044
20.0	5.04	0.045
20.5	5.21	0.046
21.0	5.37	0.046
21.5	5.54	0.047
22.0	5.71	0.048
22.5	5.88	0.048
23.0	6.06	0.049
23.5	6.23	0.050
24.0	6.42	0.051
24.5	6.60	0.051
25.0	6.79	0.052

電流値	NO ₂ 濃度	NO ₂ 濃度
μA	μg	ppm
25.5	6.98	0.053
26.0	7.17	0.054
26.5	7.36	0.054
27.0	7.56	0.055
27.5	7.77	0.056
28.0	7.97	0.056
28.5	8.18	0.057
29.0	8.39	0.058
29.5	8.61	0.059
30.0	8.83	0.059
30.5	9.05	0.060
31.0	9.28	0.061
31.5	9.51	0.062
32.0	9.74	0.062
32.5	9.98	0.063
33.0	10.22	0.064
33.5	10.46	0.065
34.0	10.71	0.065
34.5	10.96	0.066
35.0	11.22	0.067
35.5	11.48	0.068
36.0	11.74	0.069
36.5	12.01	0.069
37.0	12.29	0.070
37.5	12.56	0.071
38.0	12.84	0.072
38.5	13.13	0.072
39.0	13.42	0.073
39.5	13.72	0.074
40.0	14.01	0.075
40.5	14.32	0.076
41.0	14.63	0.076
41.5	14.94	0.077
42.0	15.26	0.078
42.5	15.58	0.079
43.0	15.91	0.080
43.5	16.24	0.081
44.0	16.58	0.081
44.5	16.92	0.082
45.0	17.27	0.083
45.5	17.62	0.084
46.0	17.97	0.085
46.5	18.34	0.086
47.0	18.70	0.086
47.5	19.08	0.087
48.0	19.46	0.088
48.5	19.84	0.089
49.0	20.23	0.090
49.5	20.62	0.091
50.0	21.02	0.092