

八王子
2010年12月
検量線NO₂
館町0402

係数
0.0210

定数M3
6.9990E-05

定数M2
7.1235E-04

定数M1
0.20988821

電流値 μA	NO ₂ 濃度 μg	NO ₂ 濃度 ppm
0.0	0.00	0.000
0.5	0.11	0.007
1.0	0.21	0.010
1.5	0.32	0.012
2.0	0.42	0.014
2.5	0.53	0.015
3.0	0.64	0.017
3.5	0.75	0.018
4.0	0.86	0.019
4.5	0.97	0.021
5.0	1.08	0.022
5.5	1.19	0.023
6.0	1.30	0.024
6.5	1.41	0.025
7.0	1.53	0.026
7.5	1.64	0.027
8.0	1.76	0.028
8.5	1.88	0.029
9.0	2.00	0.030
9.5	2.12	0.031
10.0	2.24	0.031
10.5	2.36	0.032
11.0	2.49	0.033
11.5	2.61	0.034
12.0	2.74	0.035
12.5	2.87	0.036
13.0	3.00	0.036
13.5	3.14	0.037
14.0	3.27	0.038
14.5	3.41	0.039
15.0	3.54	0.040
15.5	3.69	0.040
16.0	3.83	0.041
16.5	3.97	0.042
17.0	4.12	0.043
17.5	4.27	0.043
18.0	4.42	0.044
18.5	4.57	0.045
19.0	4.73	0.046
19.5	4.88	0.046
20.0	5.04	0.047
20.5	5.21	0.048
21.0	5.37	0.049
21.5	5.54	0.049
22.0	5.71	0.050
22.5	5.88	0.051
23.0	6.06	0.052
23.5	6.23	0.052
24.0	6.42	0.053
24.5	6.60	0.054
25.0	6.79	0.055

電流値 μA	NO ₂ 濃度 μg	NO ₂ 濃度 ppm
25.5	6.98	0.055
26.0	7.17	0.056
26.5	7.36	0.057
27.0	7.56	0.058
27.5	7.77	0.059
28.0	7.97	0.059
28.5	8.18	0.060
29.0	8.39	0.061
29.5	8.61	0.062
30.0	8.83	0.062
30.5	9.05	0.063
31.0	9.28	0.064
31.5	9.51	0.065
32.0	9.74	0.066
32.5	9.98	0.066
33.0	10.22	0.067
33.5	10.46	0.068
34.0	10.71	0.069
34.5	10.96	0.070
35.0	11.22	0.070
35.5	11.48	0.071
36.0	11.74	0.072
36.5	12.01	0.073
37.0	12.29	0.074
37.5	12.56	0.074
38.0	12.84	0.075
38.5	13.13	0.076
39.0	13.42	0.077
39.5	13.72	0.078
40.0	14.01	0.079
40.5	14.32	0.079
41.0	14.63	0.080
41.5	14.94	0.081
42.0	15.26	0.082
42.5	15.58	0.083
43.0	15.91	0.084
43.5	16.24	0.085
44.0	16.58	0.086
44.5	16.92	0.086
45.0	17.27	0.087
45.5	17.62	0.088
46.0	17.97	0.089
46.5	18.34	0.090
47.0	18.70	0.091
47.5	19.08	0.092
48.0	19.46	0.093
48.5	19.84	0.094
49.0	20.23	0.094
49.5	20.62	0.095
50.0	21.02	0.096