

八王子
2011年12月
検量線NO₂
館町0402

係数 0.0174
定数M3 7.4450E-05
定数M2 1.2796E-04
定数M1 0.20346651

電流値 μA	NO ₂ 濃度 μg	NO ₂ 濃度 ppm
0.0	0.00	0.000
0.5	0.10	0.006
1.0	0.20	0.008
1.5	0.31	0.010
2.0	0.41	0.011
2.5	0.51	0.012
3.0	0.61	0.014
3.5	0.72	0.015
4.0	0.82	0.016
4.5	0.92	0.017
5.0	1.03	0.018
5.5	1.14	0.019
6.0	1.24	0.019
6.5	1.35	0.020
7.0	1.46	0.021
7.5	1.56	0.022
8.0	1.67	0.022
8.5	1.78	0.023
9.0	1.90	0.024
9.5	2.01	0.025
10.0	2.12	0.025
10.5	2.24	0.026
11.0	2.35	0.027
11.5	2.47	0.027
12.0	2.59	0.028
12.5	2.71	0.029
13.0	2.83	0.029
13.5	2.95	0.030
14.0	3.08	0.031
14.5	3.20	0.031
15.0	3.33	0.032
15.5	3.46	0.032
16.0	3.59	0.033
16.5	3.73	0.034
17.0	3.86	0.034
17.5	4.00	0.035
18.0	4.14	0.035
18.5	4.28	0.036
19.0	4.42	0.037
19.5	4.57	0.037
20.0	4.72	0.038
20.5	4.87	0.038
21.0	5.02	0.039
21.5	5.17	0.040
22.0	5.33	0.040
22.5	5.49	0.041
23.0	5.65	0.041
23.5	5.82	0.042
24.0	5.99	0.043
24.5	6.16	0.043
25.0	6.33	0.044

電流値 μA	NO ₂ 濃度 μg	NO ₂ 濃度 ppm
25.5	6.51	0.044
26.0	6.69	0.045
26.5	6.87	0.046
27.0	7.05	0.046
27.5	7.24	0.047
28.0	7.43	0.047
28.5	7.63	0.048
29.0	7.82	0.049
29.5	8.02	0.049
30.0	8.23	0.050
30.5	8.44	0.051
31.0	8.65	0.051
31.5	8.86	0.052
32.0	9.08	0.052
32.5	9.30	0.053
33.0	9.53	0.054
33.5	9.76	0.054
34.0	9.99	0.055
34.5	10.23	0.056
35.0	10.47	0.056
35.5	10.72	0.057
36.0	10.96	0.058
36.5	11.22	0.058
37.0	11.47	0.059
37.5	11.74	0.060
38.0	12.00	0.060
38.5	12.27	0.061
39.0	12.55	0.062
39.5	12.82	0.062
40.0	13.11	0.063
40.5	13.40	0.064
41.0	13.69	0.064
41.5	13.99	0.065
42.0	14.29	0.066
42.5	14.59	0.066
43.0	14.90	0.067
43.5	15.22	0.068
44.0	15.54	0.069
44.5	15.87	0.069
45.0	16.20	0.070
45.5	16.54	0.071
46.0	16.88	0.071
46.5	17.22	0.072
47.0	17.58	0.073
47.5	17.93	0.074
48.0	18.29	0.074
48.5	18.66	0.075
49.0	19.04	0.076
49.5	19.41	0.077
50.0	19.80	0.077