

八王子
2012年12月

検量線NO₂
館町0402

係数
0.0210

定数M3
9.4100E-05

定数M2
-8.0650E-05

定数M1
0.21543172

電流値	NO ₂ 濃度	NO ₂ 濃度
μA	μg	ppm
0.0	0.00	0.000
0.5	0.11	0.007
1.0	0.22	0.010
1.5	0.32	0.012
2.0	0.43	0.014
2.5	0.54	0.015
3.0	0.65	0.017
3.5	0.76	0.018
4.0	0.87	0.020
4.5	0.98	0.021
5.0	1.09	0.022
5.5	1.20	0.023
6.0	1.31	0.024
6.5	1.42	0.025
7.0	1.54	0.026
7.5	1.65	0.027
8.0	1.77	0.028
8.5	1.88	0.029
9.0	2.00	0.030
9.5	2.12	0.031
10.0	2.24	0.031
10.5	2.36	0.032
11.0	2.49	0.033
11.5	2.61	0.034
12.0	2.74	0.035
12.5	2.86	0.036
13.0	2.99	0.036
13.5	3.13	0.037
14.0	3.26	0.038
14.5	3.39	0.039
15.0	3.53	0.039
15.5	3.67	0.040
16.0	3.81	0.041
16.5	3.96	0.042
17.0	4.10	0.043
17.5	4.25	0.043
18.0	4.40	0.044
18.5	4.55	0.045
19.0	4.71	0.046
19.5	4.87	0.046
20.0	5.03	0.047
20.5	5.19	0.048
21.0	5.36	0.049
21.5	5.53	0.049
22.0	5.70	0.050
22.5	5.88	0.051
23.0	6.06	0.052
23.5	6.24	0.052
24.0	6.42	0.053
24.5	6.61	0.054
25.0	6.81	0.055

電流値	NO ₂ 濃度	NO ₂ 濃度
μA	μg	ppm
25.5	7.00	0.056
26.0	7.20	0.056
26.5	7.40	0.057
27.0	7.61	0.058
27.5	7.82	0.059
28.0	8.03	0.060
28.5	8.25	0.060
29.0	8.47	0.061
29.5	8.70	0.062
30.0	8.93	0.063
30.5	9.17	0.064
31.0	9.40	0.064
31.5	9.65	0.065
32.0	9.89	0.066
32.5	10.15	0.067
33.0	10.40	0.068
33.5	10.66	0.069
34.0	10.93	0.069
34.5	11.20	0.070
35.0	11.48	0.071
35.5	11.76	0.072
36.0	12.04	0.073
36.5	12.33	0.074
37.0	12.63	0.075
37.5	12.93	0.076
38.0	13.23	0.076
38.5	13.54	0.077
39.0	13.86	0.078
39.5	14.18	0.079
40.0	14.51	0.080
40.5	14.84	0.081
41.0	15.18	0.082
41.5	15.53	0.083
42.0	15.88	0.084
42.5	16.23	0.085
43.0	16.60	0.086
43.5	16.96	0.086
44.0	17.34	0.087
44.5	17.72	0.088
45.0	18.11	0.089
45.5	18.50	0.090
46.0	18.90	0.091
46.5	19.30	0.092
47.0	19.72	0.093
47.5	20.14	0.094
48.0	20.56	0.095
48.5	20.99	0.096
49.0	21.43	0.097
49.5	21.88	0.098
50.0	22.33	0.099