

八王子
2014年6月

検量線NO₂
館町0402

係数
0.0190

定数M3
8.8310E-05

定数M2
-1.7332E-04

定数M1
0.22578897

電流値 μA	NO ₂ 濃度 μg	NO ₂ 濃度 ppm
0.0	0.00	0.000
0.5	0.11	0.006
1.0	0.23	0.009
1.5	0.34	0.011
2.0	0.45	0.013
2.5	0.56	0.014
3.0	0.68	0.016
3.5	0.79	0.017
4.0	0.91	0.018
4.5	1.02	0.019
5.0	1.14	0.020
5.5	1.25	0.021
6.0	1.37	0.022
6.5	1.48	0.023
7.0	1.60	0.024
7.5	1.72	0.025
8.0	1.84	0.026
8.5	1.96	0.027
9.0	2.08	0.027
9.5	2.21	0.028
10.0	2.33	0.029
10.5	2.45	0.030
11.0	2.58	0.031
11.5	2.71	0.031
12.0	2.84	0.032
12.5	2.97	0.033
13.0	3.10	0.033
13.5	3.23	0.034
14.0	3.37	0.035
14.5	3.51	0.036
15.0	3.65	0.036
15.5	3.79	0.037
16.0	3.93	0.038
16.5	4.08	0.038
17.0	4.22	0.039
17.5	4.37	0.040
18.0	4.52	0.040
18.5	4.68	0.041
19.0	4.83	0.042
19.5	4.99	0.042
20.0	5.15	0.043
20.5	5.32	0.044
21.0	5.48	0.044
21.5	5.65	0.045
22.0	5.82	0.046
22.5	6.00	0.047
23.0	6.18	0.047
23.5	6.36	0.048
24.0	6.54	0.049
24.5	6.73	0.049
25.0	6.92	0.050

電流値 μA	NO ₂ 濃度 μg	NO ₂ 濃度 ppm
25.5	7.11	0.051
26.0	7.31	0.051
26.5	7.51	0.052
27.0	7.71	0.053
27.5	7.91	0.053
28.0	8.12	0.054
28.5	8.34	0.055
29.0	8.56	0.056
29.5	8.78	0.056
30.0	9.00	0.057
30.5	9.23	0.058
31.0	9.46	0.058
31.5	9.70	0.059
32.0	9.94	0.060
32.5	10.19	0.061
33.0	10.44	0.061
33.5	10.69	0.062
34.0	10.95	0.063
34.5	11.21	0.064
35.0	11.48	0.064
35.5	11.75	0.065
36.0	12.02	0.066
36.5	12.30	0.067
37.0	12.59	0.067
37.5	12.88	0.068
38.0	13.18	0.069
38.5	13.48	0.070
39.0	13.78	0.071
39.5	14.09	0.071
40.0	14.41	0.072
40.5	14.73	0.073
41.0	15.05	0.074
41.5	15.38	0.075
42.0	15.72	0.075
42.5	16.06	0.076
43.0	16.41	0.077
43.5	16.76	0.078
44.0	17.12	0.079
44.5	17.49	0.079
45.0	17.86	0.080
45.5	18.23	0.081
46.0	18.62	0.082
46.5	19.00	0.083
47.0	19.40	0.084
47.5	19.80	0.085
48.0	20.20	0.085
48.5	20.62	0.086
49.0	21.04	0.087
49.5	21.46	0.088
50.0	21.89	0.089