

八王子  
2015年12月

検量線NO<sub>2</sub>  
館町0402

係数  
0.0220

定数M3  
2.9340E-05

定数M2  
2.9128E-03

定数M1  
0.17750918

電流値	NO <sub>2</sub> 濃度	NO <sub>2</sub> 濃度
μA	μg	ppm
0.0	0.00	0.000
0.5	0.09	0.007
1.0	0.18	0.009
1.5	0.27	0.011
2.0	0.37	0.013
2.5	0.46	0.015
3.0	0.56	0.016
3.5	0.66	0.018
4.0	0.76	0.019
4.5	0.86	0.020
5.0	0.96	0.022
5.5	1.07	0.023
6.0	1.18	0.024
6.5	1.28	0.025
7.0	1.40	0.026
7.5	1.51	0.027
8.0	1.62	0.028
8.5	1.74	0.029
9.0	1.85	0.030
9.5	1.97	0.031
10.0	2.10	0.032
10.5	2.22	0.033
11.0	2.34	0.034
11.5	2.47	0.035
12.0	2.60	0.035
12.5	2.73	0.036
13.0	2.86	0.037
13.5	3.00	0.038
14.0	3.14	0.039
14.5	3.28	0.040
15.0	3.42	0.041
15.5	3.56	0.042
16.0	3.71	0.042
16.5	3.85	0.043
17.0	4.00	0.044
17.5	4.16	0.045
18.0	4.31	0.046
18.5	4.47	0.047
19.0	4.63	0.047
19.5	4.79	0.048
20.0	4.95	0.049
20.5	5.12	0.050
21.0	5.28	0.051
21.5	5.45	0.051
22.0	5.63	0.052
22.5	5.80	0.053
23.0	5.98	0.054
23.5	6.16	0.055
24.0	6.34	0.055
24.5	6.53	0.056
25.0	6.72	0.057

電流値	NO <sub>2</sub> 濃度	NO <sub>2</sub> 濃度
μA	μg	ppm
25.5	6.91	0.058
26.0	7.10	0.059
26.5	7.30	0.059
27.0	7.49	0.060
27.5	7.69	0.061
28.0	7.90	0.062
28.5	8.10	0.063
29.0	8.31	0.063
29.5	8.52	0.064
30.0	8.74	0.065
30.5	8.96	0.066
31.0	9.18	0.067
31.5	9.40	0.067
32.0	9.62	0.068
32.5	9.85	0.069
33.0	10.08	0.070
33.5	10.32	0.071
34.0	10.56	0.071
34.5	10.80	0.072
35.0	11.04	0.073
35.5	11.29	0.074
36.0	11.53	0.075
36.5	11.79	0.076
37.0	12.04	0.076
37.5	12.30	0.077
38.0	12.56	0.078
38.5	12.83	0.079
39.0	13.09	0.080
39.5	13.36	0.080
40.0	13.64	0.081
40.5	13.92	0.082
41.0	14.20	0.083
41.5	14.48	0.084
42.0	14.77	0.085
42.5	15.06	0.085
43.0	15.35	0.086
43.5	15.65	0.087
44.0	15.95	0.088
44.5	16.25	0.089
45.0	16.56	0.090
45.5	16.87	0.090
46.0	17.18	0.091
46.5	17.50	0.092
47.0	17.82	0.093
47.5	18.15	0.094
48.0	18.48	0.095
48.5	18.81	0.095
49.0	19.14	0.096
49.5	19.48	0.097
50.0	19.82	0.098