

八王子
2016年5月
検量線NO₂
館町0402

係数
0.0190

定数M3
3.1300E-06

定数M2
4.2733E-03

定数M1
0.16269724

電流値 μA	NO ₂ 濃度 μg	NO ₂ 濃度 ppm
0.0	0.00	0.000
0.5	0.08	0.005
1.0	0.17	0.008
1.5	0.25	0.010
2.0	0.34	0.011
2.5	0.43	0.012
3.0	0.53	0.014
3.5	0.62	0.015
4.0	0.72	0.016
4.5	0.82	0.017
5.0	0.92	0.018
5.5	1.02	0.019
6.0	1.13	0.020
6.5	1.24	0.021
7.0	1.35	0.022
7.5	1.46	0.023
8.0	1.58	0.024
8.5	1.69	0.025
9.0	1.81	0.026
9.5	1.93	0.026
10.0	2.06	0.027
10.5	2.18	0.028
11.0	2.31	0.029
11.5	2.44	0.030
12.0	2.57	0.030
12.5	2.71	0.031
13.0	2.84	0.032
13.5	2.98	0.033
14.0	3.12	0.034
14.5	3.27	0.034
15.0	3.41	0.035
15.5	3.56	0.036
16.0	3.71	0.037
16.5	3.86	0.037
17.0	4.02	0.038
17.5	4.17	0.039
18.0	4.33	0.040
18.5	4.49	0.040
19.0	4.66	0.041
19.5	4.82	0.042
20.0	4.99	0.042
20.5	5.16	0.043
21.0	5.33	0.044
21.5	5.50	0.045
22.0	5.68	0.045
22.5	5.86	0.046
23.0	6.04	0.047
23.5	6.22	0.047
24.0	6.41	0.048
24.5	6.60	0.049
25.0	6.79	0.050

電流値 μA	NO ₂ 濃度 μg	NO ₂ 濃度 ppm
25.5	6.98	0.050
26.0	7.17	0.051
26.5	7.37	0.052
27.0	7.57	0.052
27.5	7.77	0.053
28.0	7.97	0.054
28.5	8.18	0.054
29.0	8.39	0.055
29.5	8.60	0.056
30.0	8.81	0.056
30.5	9.03	0.057
31.0	9.24	0.058
31.5	9.46	0.058
32.0	9.68	0.059
32.5	9.91	0.060
33.0	10.14	0.061
33.5	10.36	0.061
34.0	10.59	0.062
34.5	10.83	0.063
35.0	11.06	0.063
35.5	11.30	0.064
36.0	11.54	0.065
36.5	11.78	0.065
37.0	12.03	0.066
37.5	12.28	0.067
38.0	12.52	0.067
38.5	12.78	0.068
39.0	13.03	0.069
39.5	13.29	0.069
40.0	13.55	0.070
40.5	13.81	0.071
41.0	14.07	0.071
41.5	14.34	0.072
42.0	14.60	0.073
42.5	14.87	0.073
43.0	15.15	0.074
43.5	15.42	0.075
44.0	15.70	0.075
44.5	15.98	0.076
45.0	16.26	0.077
45.5	16.54	0.077
46.0	16.83	0.078
46.5	17.12	0.079
47.0	17.41	0.079
47.5	17.71	0.080
48.0	18.00	0.081
48.5	18.30	0.081
49.0	18.60	0.082
49.5	18.90	0.083
50.0	19.21	0.083