

八王子
2012年12月

検量線NO₂
西南0508

係数
0.0210

定数M3
1.0695E-04

定数M2
-1.2543E-03

定数M1
0.22513172

電流値 μA	NO ₂ 濃度 μg	NO ₂ 濃度 ppm
0.0	0.00	0.000
0.5	0.11	0.007
1.0	0.22	0.010
1.5	0.34	0.012
2.0	0.45	0.014
2.5	0.56	0.016
3.0	0.67	0.017
3.5	0.78	0.019
4.0	0.89	0.020
4.5	1.00	0.021
5.0	1.11	0.022
5.5	1.22	0.023
6.0	1.33	0.024
6.5	1.44	0.025
7.0	1.55	0.026
7.5	1.66	0.027
8.0	1.78	0.028
8.5	1.89	0.029
9.0	2.00	0.030
9.5	2.12	0.031
10.0	2.23	0.031
10.5	2.35	0.032
11.0	2.47	0.033
11.5	2.59	0.034
12.0	2.71	0.035
12.5	2.83	0.035
13.0	2.95	0.036
13.5	3.07	0.037
14.0	3.20	0.038
14.5	3.33	0.038
15.0	3.46	0.039
15.5	3.59	0.040
16.0	3.72	0.041
16.5	3.85	0.041
17.0	3.99	0.042
17.5	4.13	0.043
18.0	4.27	0.043
18.5	4.41	0.044
19.0	4.56	0.045
19.5	4.71	0.046
20.0	4.86	0.046
20.5	5.01	0.047
21.0	5.17	0.048
21.5	5.32	0.048
22.0	5.48	0.049
22.5	5.65	0.050
23.0	5.82	0.051
23.5	5.99	0.051
24.0	6.16	0.052
24.5	6.34	0.053
25.0	6.52	0.054

電流値 μA	NO ₂ 濃度 μg	NO ₂ 濃度 ppm
25.5	6.70	0.054
26.0	6.89	0.055
26.5	7.08	0.056
27.0	7.27	0.057
27.5	7.47	0.057
28.0	7.67	0.058
28.5	7.87	0.059
29.0	8.08	0.060
29.5	8.30	0.061
30.0	8.51	0.061
30.5	8.73	0.062
31.0	8.96	0.063
31.5	9.19	0.064
32.0	9.42	0.064
32.5	9.66	0.065
33.0	9.91	0.066
33.5	10.16	0.067
34.0	10.41	0.068
34.5	10.67	0.069
35.0	10.93	0.069
35.5	11.20	0.070
36.0	11.47	0.071
36.5	11.75	0.072
37.0	12.03	0.073
37.5	12.32	0.074
38.0	12.61	0.075
38.5	12.91	0.075
39.0	13.22	0.076
39.5	13.53	0.077
40.0	13.84	0.078
40.5	14.17	0.079
41.0	14.49	0.080
41.5	14.83	0.081
42.0	15.17	0.082
42.5	15.51	0.083
43.0	15.86	0.084
43.5	16.22	0.085
44.0	16.59	0.086
44.5	16.96	0.086
45.0	17.34	0.087
45.5	17.72	0.088
46.0	18.11	0.089
46.5	18.51	0.090
47.0	18.91	0.091
47.5	19.33	0.092
48.0	19.74	0.093
48.5	20.17	0.094
49.0	20.60	0.095
49.5	21.04	0.096
50.0	21.49	0.097