

八王子
2013年12月

検量線NO₂
西南0508

係数
0.0290

定数M3
1.5590E-04

定数M2
-2.8629E-03

定数M1
0.24978641

電流値	NO ₂ 濃度	NO ₂ 濃度
μA	μg	ppm
0.0	0.00	0.000
0.5	0.12	0.010
1.0	0.25	0.015
1.5	0.37	0.018
2.0	0.49	0.020
2.5	0.61	0.023
3.0	0.73	0.025
3.5	0.85	0.027
4.0	0.96	0.028
4.5	1.08	0.030
5.0	1.20	0.032
5.5	1.31	0.033
6.0	1.43	0.035
6.5	1.55	0.036
7.0	1.66	0.037
7.5	1.78	0.039
8.0	1.89	0.040
8.5	2.01	0.041
9.0	2.13	0.042
9.5	2.25	0.044
10.0	2.37	0.045
10.5	2.49	0.046
11.0	2.61	0.047
11.5	2.73	0.048
12.0	2.85	0.049
12.5	2.98	0.050
13.0	3.11	0.051
13.5	3.23	0.052
14.0	3.36	0.053
14.5	3.50	0.054
15.0	3.63	0.055
15.5	3.76	0.056
16.0	3.90	0.057
16.5	4.04	0.058
17.0	4.18	0.059
17.5	4.33	0.060
18.0	4.48	0.061
18.5	4.63	0.062
19.0	4.78	0.063
19.5	4.94	0.064
20.0	5.10	0.065
20.5	5.26	0.067
21.0	5.43	0.068
21.5	5.60	0.069
22.0	5.77	0.070
22.5	5.95	0.071
23.0	6.13	0.072
23.5	6.31	0.073
24.0	6.50	0.074
24.5	6.69	0.075
25.0	6.89	0.076

電流値	NO ₂ 濃度	NO ₂ 濃度
μA	μg	ppm
25.5	7.09	0.077
26.0	7.30	0.078
26.5	7.51	0.079
27.0	7.73	0.081
27.5	7.95	0.082
28.0	8.17	0.083
28.5	8.40	0.084
29.0	8.64	0.085
29.5	8.88	0.086
30.0	9.13	0.088
30.5	9.38	0.089
31.0	9.64	0.090
31.5	9.90	0.091
32.0	10.17	0.092
32.5	10.45	0.094
33.0	10.73	0.095
33.5	11.02	0.096
34.0	11.31	0.098
34.5	11.61	0.099
35.0	11.92	0.100
35.5	12.23	0.101
36.0	12.56	0.103
36.5	12.88	0.104
37.0	13.22	0.105
37.5	13.56	0.107
38.0	13.91	0.108
38.5	14.27	0.110
39.0	14.64	0.111
39.5	15.01	0.112
40.0	15.39	0.114
40.5	15.78	0.115
41.0	16.17	0.117
41.5	16.58	0.118
42.0	16.99	0.120
42.5	17.41	0.121
43.0	17.84	0.122
43.5	18.28	0.124
44.0	18.73	0.126
44.5	19.18	0.127
45.0	19.65	0.129
45.5	20.12	0.130
46.0	20.61	0.132
46.5	21.10	0.133
47.0	21.60	0.135
47.5	22.11	0.136
48.0	22.63	0.138
48.5	23.17	0.140
49.0	23.71	0.141
49.5	24.26	0.143
50.0	24.82	0.144