

八王子  
2014年6月

検量線NO<sub>2</sub>  
西南0508

係数  
0.0190

定数M3  
1.0307E-04

定数M2  
-9.1923E-04

定数M1  
0.22381847

電流値 μA	NO <sub>2</sub> 濃度 μg	NO <sub>2</sub> 濃度 ppm
0.0	0.00	0.000
0.5	0.11	0.006
1.0	0.22	0.009
1.5	0.33	0.011
2.0	0.44	0.013
2.5	0.56	0.014
3.0	0.67	0.016
3.5	0.78	0.017
4.0	0.89	0.018
4.5	1.00	0.019
5.0	1.11	0.020
5.5	1.22	0.021
6.0	1.33	0.022
6.5	1.44	0.023
7.0	1.56	0.024
7.5	1.67	0.025
8.0	1.78	0.025
8.5	1.90	0.026
9.0	2.02	0.027
9.5	2.13	0.028
10.0	2.25	0.029
10.5	2.37	0.029
11.0	2.49	0.030
11.5	2.61	0.031
12.0	2.73	0.031
12.5	2.86	0.032
13.0	2.98	0.033
13.5	3.11	0.034
14.0	3.24	0.034
14.5	3.37	0.035
15.0	3.50	0.036
15.5	3.63	0.036
16.0	3.77	0.037
16.5	3.91	0.038
17.0	4.05	0.038
17.5	4.19	0.039
18.0	4.33	0.040
18.5	4.48	0.040
19.0	4.63	0.041
19.5	4.78	0.042
20.0	4.93	0.042
20.5	5.09	0.043
21.0	5.25	0.044
21.5	5.41	0.044
22.0	5.58	0.045
22.5	5.74	0.046
23.0	5.92	0.046
23.5	6.09	0.047
24.0	6.27	0.048
24.5	6.45	0.048
25.0	6.63	0.049

電流値 μA	NO <sub>2</sub> 濃度 μg	NO <sub>2</sub> 濃度 ppm
25.5	6.82	0.050
26.0	7.01	0.050
26.5	7.20	0.051
27.0	7.40	0.052
27.5	7.60	0.052
28.0	7.81	0.053
28.5	8.02	0.054
29.0	8.23	0.055
29.5	8.45	0.055
30.0	8.67	0.056
30.5	8.90	0.057
31.0	9.13	0.057
31.5	9.36	0.058
32.0	9.60	0.059
32.5	9.84	0.060
33.0	10.09	0.060
33.5	10.34	0.061
34.0	10.60	0.062
34.5	10.86	0.063
35.0	11.13	0.063
35.5	11.40	0.064
36.0	11.67	0.065
36.5	11.96	0.066
37.0	12.24	0.066
37.5	12.54	0.067
38.0	12.83	0.068
38.5	13.14	0.069
39.0	13.44	0.070
39.5	13.76	0.070
40.0	14.08	0.071
40.5	14.40	0.072
41.0	14.74	0.073
41.5	15.07	0.074
42.0	15.42	0.075
42.5	15.76	0.075
43.0	16.12	0.076
43.5	16.48	0.077
44.0	16.85	0.078
44.5	17.22	0.079
45.0	17.60	0.080
45.5	17.99	0.081
46.0	18.38	0.081
46.5	18.78	0.082
47.0	19.19	0.083
47.5	19.60	0.084
48.0	20.02	0.085
48.5	20.45	0.086
49.0	20.89	0.087
49.5	21.33	0.088
50.0	21.78	0.089