

八王子  
2015年12月  
検量線NO<sub>2</sub>  
西南0508

係数  
0.0220

定数M3  
7.1000E-05

定数M2  
6.9930E-04

定数M1  
0.19434784

電流値	NO <sub>2</sub> 濃度	NO <sub>2</sub> 濃度
μA	μg	ppm
0.0	0.00	0.000
0.5	0.10	0.007
1.0	0.20	0.010
1.5	0.29	0.012
2.0	0.39	0.014
2.5	0.49	0.015
3.0	0.59	0.017
3.5	0.69	0.018
4.0	0.79	0.020
4.5	0.90	0.021
5.0	1.00	0.022
5.5	1.10	0.023
6.0	1.21	0.024
6.5	1.31	0.025
7.0	1.42	0.026
7.5	1.53	0.027
8.0	1.64	0.028
8.5	1.75	0.029
9.0	1.86	0.030
9.5	1.97	0.031
10.0	2.08	0.032
10.5	2.20	0.033
11.0	2.32	0.034
11.5	2.44	0.034
12.0	2.56	0.035
12.5	2.68	0.036
13.0	2.80	0.037
13.5	2.93	0.038
14.0	3.05	0.038
14.5	3.18	0.039
15.0	3.31	0.040
15.5	3.44	0.041
16.0	3.58	0.042
16.5	3.72	0.042
17.0	3.85	0.043
17.5	4.00	0.044
18.0	4.14	0.045
18.5	4.28	0.046
19.0	4.43	0.046
19.5	4.58	0.047
20.0	4.73	0.048
20.5	4.89	0.049
21.0	5.05	0.049
21.5	5.21	0.050
22.0	5.37	0.051
22.5	5.54	0.052
23.0	5.70	0.053
23.5	5.87	0.053
24.0	6.05	0.054
24.5	6.23	0.055
25.0	6.41	0.056

電流値	NO <sub>2</sub> 濃度	NO <sub>2</sub> 濃度
μA	μg	ppm
25.5	6.59	0.056
26.0	6.77	0.057
26.5	6.96	0.058
27.0	7.15	0.059
27.5	7.35	0.060
28.0	7.55	0.060
28.5	7.75	0.061
29.0	7.96	0.062
29.5	8.16	0.063
30.0	8.38	0.064
30.5	8.59	0.064
31.0	8.81	0.065
31.5	9.04	0.066
32.0	9.26	0.067
32.5	9.49	0.068
33.0	9.73	0.069
33.5	9.96	0.069
34.0	10.21	0.070
34.5	10.45	0.071
35.0	10.70	0.072
35.5	10.96	0.073
36.0	11.22	0.074
36.5	11.48	0.075
37.0	11.74	0.075
37.5	12.02	0.076
38.0	12.29	0.077
38.5	12.57	0.078
39.0	12.85	0.079
39.5	13.14	0.080
40.0	13.44	0.081
40.5	13.73	0.082
41.0	14.04	0.082
41.5	14.34	0.083
42.0	14.66	0.084
42.5	14.97	0.085
43.0	15.29	0.086
43.5	15.62	0.087
44.0	15.95	0.088
44.5	16.29	0.089
45.0	16.63	0.090
45.5	16.98	0.091
46.0	17.33	0.092
46.5	17.69	0.093
47.0	18.05	0.093
47.5	18.42	0.094
48.0	18.79	0.095
48.5	19.17	0.096
49.0	19.56	0.097
49.5	19.95	0.098
50.0	20.34	0.099