

八王子
2016年5月

検量線NO₂
西南0508

係数
0.0190

定数M3
5.5480E-05

定数M2
1.3049E-03

定数M1
0.20702598

電流値 μA	NO ₂ 濃度 μg	NO ₂ 濃度 ppm
0.0	0.00	0.000
0.5	0.10	0.006
1.0	0.21	0.009
1.5	0.31	0.011
2.0	0.42	0.012
2.5	0.53	0.014
3.0	0.63	0.015
3.5	0.74	0.016
4.0	0.85	0.018
4.5	0.96	0.019
5.0	1.07	0.020
5.5	1.19	0.021
6.0	1.30	0.022
6.5	1.42	0.023
7.0	1.53	0.024
7.5	1.65	0.024
8.0	1.77	0.025
8.5	1.89	0.026
9.0	2.01	0.027
9.5	2.13	0.028
10.0	2.26	0.029
10.5	2.38	0.029
11.0	2.51	0.030
11.5	2.64	0.031
12.0	2.77	0.032
12.5	2.90	0.032
13.0	3.03	0.033
13.5	3.17	0.034
14.0	3.31	0.035
14.5	3.45	0.035
15.0	3.59	0.036
15.5	3.73	0.037
16.0	3.87	0.037
16.5	4.02	0.038
17.0	4.17	0.039
17.5	4.32	0.039
18.0	4.47	0.040
18.5	4.63	0.041
19.0	4.79	0.042
19.5	4.94	0.042
20.0	5.11	0.043
20.5	5.27	0.044
21.0	5.44	0.044
21.5	5.61	0.045
22.0	5.78	0.046
22.5	5.95	0.046
23.0	6.13	0.047
23.5	6.31	0.048
24.0	6.49	0.048
24.5	6.67	0.049
25.0	6.86	0.050

電流値 μA	NO ₂ 濃度 μg	NO ₂ 濃度 ppm
25.5	7.05	0.050
26.0	7.24	0.051
26.5	7.43	0.052
27.0	7.63	0.052
27.5	7.83	0.053
28.0	8.04	0.054
28.5	8.24	0.055
29.0	8.45	0.055
29.5	8.67	0.056
30.0	8.88	0.057
30.5	9.10	0.057
31.0	9.32	0.058
31.5	9.55	0.059
32.0	9.78	0.059
32.5	10.01	0.060
33.0	10.25	0.061
33.5	10.49	0.062
34.0	10.73	0.062
34.5	10.97	0.063
35.0	11.22	0.064
35.5	11.48	0.064
36.0	11.73	0.065
36.5	11.99	0.066
37.0	12.26	0.067
37.5	12.52	0.067
38.0	12.80	0.068
38.5	13.07	0.069
39.0	13.35	0.069
39.5	13.63	0.070
40.0	13.92	0.071
40.5	14.21	0.072
41.0	14.51	0.072
41.5	14.80	0.073
42.0	15.11	0.074
42.5	15.41	0.075
43.0	15.73	0.075
43.5	16.04	0.076
44.0	16.36	0.077
44.5	16.69	0.078
45.0	17.01	0.078
45.5	17.35	0.079
46.0	17.68	0.080
46.5	18.03	0.081
47.0	18.37	0.081
47.5	18.72	0.082
48.0	19.08	0.083
48.5	19.44	0.084
49.0	19.80	0.085
49.5	20.17	0.085
50.0	20.55	0.086