

検量線NO<sub>2</sub>  
西南0508

係数  
0.0227

定数M3  
5.5480E-05

定数M2  
1.3049E-03

定数M1  
0.20702598

八王子  
2016年12月

電流値 μA	NO <sub>2</sub> 濃度 μg	NO <sub>2</sub> 濃度 ppm
0.0	0.00	0.000
0.5	0.10	0.007
1.0	0.21	0.010
1.5	0.31	0.013
2.0	0.42	0.015
2.5	0.53	0.017
3.0	0.63	0.018
3.5	0.74	0.020
4.0	0.85	0.021
4.5	0.96	0.022
5.0	1.07	0.023
5.5	1.19	0.025
6.0	1.30	0.026
6.5	1.42	0.027
7.0	1.53	0.028
7.5	1.65	0.029
8.0	1.77	0.030
8.5	1.89	0.031
9.0	2.01	0.032
9.5	2.13	0.033
10.0	2.26	0.034
10.5	2.38	0.035
11.0	2.51	0.036
11.5	2.64	0.037
12.0	2.77	0.038
12.5	2.90	0.039
13.0	3.03	0.040
13.5	3.17	0.040
14.0	3.31	0.041
14.5	3.45	0.042
15.0	3.59	0.043
15.5	3.73	0.044
16.0	3.87	0.045
16.5	4.02	0.046
17.0	4.17	0.046
17.5	4.32	0.047
18.0	4.47	0.048
18.5	4.63	0.049
19.0	4.79	0.050
19.5	4.94	0.050
20.0	5.11	0.051
20.5	5.27	0.052
21.0	5.44	0.053
21.5	5.61	0.054
22.0	5.78	0.055
22.5	5.95	0.055
23.0	6.13	0.056
23.5	6.31	0.057
24.0	6.49	0.058
24.5	6.67	0.059
25.0	6.86	0.059

電流値 μA	NO <sub>2</sub> 濃度 μg	NO <sub>2</sub> 濃度 ppm
25.5	7.05	0.060
26.0	7.24	0.061
26.5	7.43	0.062
27.0	7.63	0.063
27.5	7.83	0.064
28.0	8.04	0.064
28.5	8.24	0.065
29.0	8.45	0.066
29.5	8.67	0.067
30.0	8.88	0.068
30.5	9.10	0.068
31.0	9.32	0.069
31.5	9.55	0.070
32.0	9.78	0.071
32.5	10.01	0.072
33.0	10.25	0.073
33.5	10.49	0.074
34.0	10.73	0.074
34.5	10.97	0.075
35.0	11.22	0.076
35.5	11.48	0.077
36.0	11.73	0.078
36.5	11.99	0.079
37.0	12.26	0.079
37.5	12.52	0.080
38.0	12.80	0.081
38.5	13.07	0.082
39.0	13.35	0.083
39.5	13.63	0.084
40.0	13.92	0.085
40.5	14.21	0.086
41.0	14.51	0.086
41.5	14.80	0.087
42.0	15.11	0.088
42.5	15.41	0.089
43.0	15.73	0.090
43.5	16.04	0.091
44.0	16.36	0.092
44.5	16.69	0.093
45.0	17.01	0.094
45.5	17.35	0.095
46.0	17.68	0.095
46.5	18.03	0.096
47.0	18.37	0.097
47.5	18.72	0.098
48.0	19.08	0.099
48.5	19.44	0.100
49.0	19.80	0.101
49.5	20.17	0.102
50.0	20.55	0.103