

八王子
2020年12月

検量線NO₂
館町0402

12月係数 0.0220

定数M3
1.0029E-04

定数M2
-3.2592E-04

定数M1
0.21429641

電流値 μA	NO ₂ 濃度 μg	NO ₂ 濃度 ppm
0.0	0.00	0.000
0.5	0.11	0.007
1.0	0.21	0.010
1.5	0.32	0.012
2.0	0.43	0.014
2.5	0.54	0.016
3.0	0.64	0.018
3.5	0.75	0.019
4.0	0.86	0.020
4.5	0.97	0.022
5.0	1.08	0.023
5.5	1.19	0.024
6.0	1.30	0.025
6.5	1.41	0.026
7.0	1.52	0.027
7.5	1.63	0.028
8.0	1.74	0.029
8.5	1.86	0.030
9.0	1.98	0.031
9.5	2.09	0.032
10.0	2.21	0.033
10.5	2.33	0.034
11.0	2.45	0.034
11.5	2.57	0.035
12.0	2.70	0.036
12.5	2.82	0.037
13.0	2.95	0.038
13.5	3.08	0.039
14.0	3.21	0.039
14.5	3.34	0.040
15.0	3.48	0.041
15.5	3.62	0.042
16.0	3.76	0.043
16.5	3.90	0.043
17.0	4.04	0.044
17.5	4.19	0.045
18.0	4.34	0.046
18.5	4.49	0.047
19.0	4.64	0.047
19.5	4.80	0.048
20.0	4.96	0.049
20.5	5.12	0.050
21.0	5.29	0.051
21.5	5.45	0.051
22.0	5.62	0.052
22.5	5.80	0.053
23.0	5.98	0.054
23.5	6.16	0.055
24.0	6.34	0.055
24.5	6.53	0.056
25.0	6.72	0.057

電流値 μA	NO ₂ 濃度 μg	NO ₂ 濃度 ppm
25.5	6.92	0.058
26.0	7.11	0.059
26.5	7.32	0.060
27.0	7.52	0.060
27.5	7.73	0.061
28.0	7.95	0.062
28.5	8.16	0.063
29.0	8.39	0.064
29.5	8.61	0.065
30.0	8.84	0.065
30.5	9.08	0.066
31.0	9.32	0.067
31.5	9.56	0.068
32.0	9.81	0.069
32.5	10.06	0.070
33.0	10.32	0.071
33.5	10.58	0.072
34.0	10.85	0.072
34.5	11.12	0.073
35.0	11.40	0.074
35.5	11.68	0.075
36.0	11.97	0.076
36.5	12.26	0.077
37.0	12.56	0.078
37.5	12.87	0.079
38.0	13.18	0.080
38.5	13.49	0.081
39.0	13.81	0.082
39.5	14.14	0.083
40.0	14.47	0.084
40.5	14.81	0.085
41.0	15.15	0.086
41.5	15.50	0.087
42.0	15.86	0.088
42.5	16.22	0.089
43.0	16.59	0.090
43.5	16.96	0.091
44.0	17.34	0.092
44.5	17.73	0.093
45.0	18.12	0.094
45.5	18.52	0.095
46.0	18.93	0.096
46.5	19.34	0.097
47.0	19.76	0.098
47.5	20.19	0.099
48.0	20.63	0.100
48.5	21.07	0.101
49.0	21.52	0.102
49.5	21.97	0.103
50.0	22.44	0.104

八王子
2020年12月

検量線NO₂
西南0508

12月係数

係数
0.0220

定数M3
9.4680E-05

定数M2
-1.2683E-04

定数M1
0.20059714

電流値 μA	NO ₂ 濃度 μg	NO ₂ 濃度 ppm
0.0	0.00	0.000
0.5	0.10	0.007
1.0	0.20	0.010
1.5	0.30	0.012
2.0	0.40	0.014
2.5	0.50	0.016
3.0	0.60	0.017
3.5	0.70	0.018
4.0	0.81	0.020
4.5	0.91	0.021
5.0	1.01	0.022
5.5	1.12	0.023
6.0	1.22	0.024
6.5	1.32	0.025
7.0	1.43	0.026
7.5	1.54	0.027
8.0	1.65	0.028
8.5	1.75	0.029
9.0	1.86	0.030
9.5	1.98	0.031
10.0	2.09	0.032
10.5	2.20	0.033
11.0	2.32	0.034
11.5	2.43	0.034
12.0	2.55	0.035
12.5	2.67	0.036
13.0	2.79	0.037
13.5	2.92	0.038
14.0	3.04	0.038
14.5	3.17	0.039
15.0	3.30	0.040
15.5	3.43	0.041
16.0	3.56	0.042
16.5	3.70	0.042
17.0	3.84	0.043
17.5	3.98	0.044
18.0	4.12	0.045
18.5	4.27	0.045
19.0	4.41	0.046
19.5	4.57	0.047
20.0	4.72	0.048
20.5	4.87	0.049
21.0	5.03	0.049
21.5	5.20	0.050
22.0	5.36	0.051
22.5	5.53	0.052
23.0	5.70	0.053
23.5	5.87	0.053
24.0	6.05	0.054
24.5	6.23	0.055
25.0	6.42	0.056

電流値 μA	NO ₂ 濃度 μg	NO ₂ 濃度 ppm
25.5	6.60	0.057
26.0	6.79	0.057
26.5	6.99	0.058
27.0	7.19	0.059
27.5	7.39	0.060
28.0	7.60	0.061
28.5	7.81	0.061
29.0	8.02	0.062
29.5	8.24	0.063
30.0	8.46	0.064
30.5	8.69	0.065
31.0	8.92	0.066
31.5	9.15	0.067
32.0	9.39	0.067
32.5	9.64	0.068
33.0	9.88	0.069
33.5	10.14	0.070
34.0	10.39	0.071
34.5	10.66	0.072
35.0	10.92	0.073
35.5	11.20	0.074
36.0	11.47	0.075
36.5	11.76	0.075
37.0	12.04	0.076
37.5	12.34	0.077
38.0	12.63	0.078
38.5	12.94	0.079
39.0	13.25	0.080
39.5	13.56	0.081
40.0	13.88	0.082
40.5	14.21	0.083
41.0	14.54	0.084
41.5	14.87	0.085
42.0	15.22	0.086
42.5	15.56	0.087
43.0	15.92	0.088
43.5	16.28	0.089
44.0	16.65	0.090
44.5	17.02	0.091
45.0	17.40	0.092
45.5	17.78	0.093
46.0	18.17	0.094
46.5	18.57	0.095
47.0	18.98	0.096
47.5	19.39	0.097
48.0	19.81	0.098
48.5	20.23	0.099
49.0	20.66	0.100
49.5	21.10	0.101
50.0	21.55	0.102

八王子ネット
2020年12月

検量線NO₂
9420

12月係数

係数
0.0220

定数M3
1.3250E-05

定数M2
1.1636E-03

定数M1
0.17180279

電流値 μA	NO ₂ 濃度 μg	NO ₂ 濃度 ppm
0.0	0.00	0.000
0.5	0.09	0.007
1.0	0.17	0.009
1.5	0.26	0.011
2.0	0.35	0.013
2.5	0.44	0.015
3.0	0.53	0.016
3.5	0.62	0.017
4.0	0.71	0.019
4.5	0.80	0.020
5.0	0.89	0.021
5.5	0.98	0.022
6.0	1.08	0.023
6.5	1.17	0.024
7.0	1.26	0.025
7.5	1.36	0.026
8.0	1.46	0.027
8.5	1.55	0.027
9.0	1.65	0.028
9.5	1.75	0.029
10.0	1.85	0.030
10.5	1.95	0.031
11.0	2.05	0.031
11.5	2.15	0.032
12.0	2.25	0.033
12.5	2.36	0.034
13.0	2.46	0.035
13.5	2.56	0.035
14.0	2.67	0.036
14.5	2.78	0.037
15.0	2.88	0.037
15.5	2.99	0.038
16.0	3.10	0.039
16.5	3.21	0.039
17.0	3.32	0.040
17.5	3.43	0.041
18.0	3.55	0.041
18.5	3.66	0.042
19.0	3.78	0.043
19.5	3.89	0.043
20.0	4.01	0.044
20.5	4.13	0.045
21.0	4.24	0.045
21.5	4.36	0.046
22.0	4.48	0.047
22.5	4.61	0.047
23.0	4.73	0.048
23.5	4.85	0.048
24.0	4.98	0.049
24.5	5.10	0.050
25.0	5.23	0.050

電流値 μA	NO ₂ 濃度 μg	NO ₂ 濃度 ppm
25.5	5.36	0.051
26.0	5.49	0.052
26.5	5.62	0.052
27.0	5.75	0.053
27.5	5.88	0.053
28.0	6.01	0.054
28.5	6.15	0.055
29.0	6.28	0.055
29.5	6.42	0.056
30.0	6.56	0.056
30.5	6.70	0.057
31.0	6.84	0.058
31.5	6.98	0.058
32.0	7.12	0.059
32.5	7.27	0.059
33.0	7.41	0.060
33.5	7.56	0.060
34.0	7.71	0.061
34.5	7.86	0.062
35.0	8.01	0.062
35.5	8.16	0.063
36.0	8.31	0.063
36.5	8.47	0.064
37.0	8.62	0.065
37.5	8.78	0.065
38.0	8.94	0.066
38.5	9.10	0.066
39.0	9.26	0.067
39.5	9.42	0.068
40.0	9.58	0.068
40.5	9.75	0.069
41.0	9.91	0.069
41.5	10.08	0.070
42.0	10.25	0.070
42.5	10.42	0.071
43.0	10.59	0.072
43.5	10.77	0.072
44.0	10.94	0.073
44.5	11.12	0.073
45.0	11.29	0.074
45.5	11.47	0.075
46.0	11.65	0.075
46.5	11.84	0.076
47.0	12.02	0.076
47.5	12.21	0.077
48.0	12.39	0.077
48.5	12.58	0.078
49.0	12.77	0.079
49.5	12.96	0.079
50.0	13.16	0.080

八王子ネット
2020年12月

検量線NO₂
9416

12月係数

係数
0.0220

定数M3
3.6600E-05

定数M2
4.3158E-04

定数M1
0.17814632

電流値 μA	NO ₂ 濃度 μg	NO ₂ 濃度 ppm
0.0	0.00	0.000
0.5	0.09	0.007
1.0	0.18	0.009
1.5	0.27	0.011
2.0	0.36	0.013
2.5	0.45	0.015
3.0	0.54	0.016
3.5	0.63	0.017
4.0	0.72	0.019
4.5	0.81	0.020
5.0	0.91	0.021
5.5	1.00	0.022
6.0	1.09	0.023
6.5	1.19	0.024
7.0	1.28	0.025
7.5	1.38	0.026
8.0	1.47	0.027
8.5	1.57	0.028
9.0	1.66	0.028
9.5	1.76	0.029
10.0	1.86	0.030
10.5	1.96	0.031
11.0	2.06	0.032
11.5	2.16	0.032
12.0	2.26	0.033
12.5	2.37	0.034
13.0	2.47	0.035
13.5	2.57	0.035
14.0	2.68	0.036
14.5	2.79	0.037
15.0	2.89	0.037
15.5	3.00	0.038
16.0	3.11	0.039
16.5	3.22	0.039
17.0	3.33	0.040
17.5	3.45	0.041
18.0	3.56	0.042
18.5	3.68	0.042
19.0	3.79	0.043
19.5	3.91	0.044
20.0	4.03	0.044
20.5	4.15	0.045
21.0	4.27	0.045
21.5	4.39	0.046
22.0	4.52	0.047
22.5	4.64	0.047
23.0	4.77	0.048
23.5	4.90	0.049
24.0	5.03	0.049
24.5	5.16	0.050
25.0	5.30	0.051

電流値 μA	NO ₂ 濃度 μg	NO ₂ 濃度 ppm
25.5	5.43	0.051
26.0	5.57	0.052
26.5	5.71	0.053
27.0	5.84	0.053
27.5	5.99	0.054
28.0	6.13	0.054
28.5	6.27	0.055
29.0	6.42	0.056
29.5	6.57	0.056
30.0	6.72	0.057
30.5	6.87	0.058
31.0	7.03	0.058
31.5	7.18	0.059
32.0	7.34	0.060
32.5	7.50	0.060
33.0	7.66	0.061
33.5	7.83	0.062
34.0	7.99	0.062
34.5	8.16	0.063
35.0	8.33	0.063
35.5	8.51	0.064
36.0	8.68	0.065
36.5	8.86	0.065
37.0	9.04	0.066
37.5	9.22	0.067
38.0	9.40	0.067
38.5	9.59	0.068
39.0	9.78	0.069
39.5	9.97	0.069
40.0	10.16	0.070
40.5	10.35	0.071
41.0	10.55	0.071
41.5	10.75	0.072
42.0	10.96	0.073
42.5	11.16	0.073
43.0	11.37	0.074
43.5	11.58	0.075
44.0	11.79	0.076
44.5	12.01	0.076
45.0	12.23	0.077
45.5	12.45	0.078
46.0	12.67	0.078
46.5	12.90	0.079
47.0	13.13	0.080
47.5	13.36	0.080
48.0	13.59	0.081
48.5	13.83	0.082
49.0	14.07	0.083
49.5	14.31	0.083
50.0	14.56	0.084