

八王子ネット  
2010年6月

検量線NO<sub>2</sub>  
9420

係数  
0.0200

定数M3  
-9.2700E-06

定数M2  
2.7871E-03

定数M1  
0.15245583

電流値 μA	NO <sub>2</sub> 濃度 μg	NO <sub>2</sub> 濃度 ppm
0.0	0.00	0.000
0.5	0.08	0.006
1.0	0.16	0.008
1.5	0.23	0.010
2.0	0.32	0.011
2.5	0.40	0.013
3.0	0.48	0.014
3.5	0.57	0.015
4.0	0.65	0.016
4.5	0.74	0.017
5.0	0.83	0.018
5.5	0.92	0.019
6.0	1.01	0.020
6.5	1.11	0.021
7.0	1.20	0.022
7.5	1.30	0.023
8.0	1.39	0.024
8.5	1.49	0.024
9.0	1.59	0.025
9.5	1.69	0.026
10.0	1.79	0.027
10.5	1.90	0.028
11.0	2.00	0.028
11.5	2.11	0.029
12.0	2.21	0.030
12.5	2.32	0.030
13.0	2.43	0.031
13.5	2.54	0.032
14.0	2.66	0.033
14.5	2.77	0.033
15.0	2.88	0.034
15.5	3.00	0.035
16.0	3.11	0.035
16.5	3.23	0.036
17.0	3.35	0.037
17.5	3.47	0.037
18.0	3.59	0.038
18.5	3.72	0.039
19.0	3.84	0.039
19.5	3.96	0.040
20.0	4.09	0.040
20.5	4.22	0.041
21.0	4.34	0.042
21.5	4.47	0.042
22.0	4.60	0.043
22.5	4.74	0.044
23.0	4.87	0.044
23.5	5.00	0.045
24.0	5.14	0.045
24.5	5.27	0.046
25.0	5.41	0.047

電流値 μA	NO <sub>2</sub> 濃度 μg	NO <sub>2</sub> 濃度 ppm
25.5	5.55	0.047
26.0	5.68	0.048
26.5	5.82	0.048
27.0	5.97	0.049
27.5	6.11	0.049
28.0	6.25	0.050
28.5	6.39	0.051
29.0	6.54	0.051
29.5	6.68	0.052
30.0	6.83	0.052
30.5	6.98	0.053
31.0	7.13	0.053
31.5	7.28	0.054
32.0	7.43	0.055
32.5	7.58	0.055
33.0	7.73	0.056
33.5	7.89	0.056
34.0	8.04	0.057
34.5	8.20	0.057
35.0	8.35	0.058
35.5	8.51	0.058
36.0	8.67	0.059
36.5	8.83	0.059
37.0	8.99	0.060
37.5	9.15	0.060
38.0	9.31	0.061
38.5	9.47	0.062
39.0	9.64	0.062
39.5	9.80	0.063
40.0	9.96	0.063
40.5	10.13	0.064
41.0	10.30	0.064
41.5	10.46	0.065
42.0	10.63	0.065
42.5	10.80	0.066
43.0	10.97	0.066
43.5	11.14	0.067
44.0	11.31	0.067
44.5	11.49	0.068
45.0	11.66	0.068
45.5	11.83	0.069
46.0	12.01	0.069
46.5	12.18	0.070
47.0	12.36	0.070
47.5	12.54	0.071
48.0	12.71	0.071
48.5	12.89	0.072
49.0	13.07	0.072
49.5	13.25	0.073
50.0	13.43	0.073