

八王子ネット
2011年6月
検量線NO₂
9420

係数 0.0250
定数M3 1.5360E-05
定数M2 1.0296E-03
定数M1 0.15794710

電流値 μA	NO ₂ 濃度 μg	NO ₂ 濃度 ppm
0.0	0.00	0.000
0.5	0.08	0.007
1.0	0.16	0.010
1.5	0.24	0.012
2.0	0.32	0.014
2.5	0.40	0.016
3.0	0.48	0.017
3.5	0.57	0.019
4.0	0.65	0.020
4.5	0.73	0.021
5.0	0.82	0.023
5.5	0.90	0.024
6.0	0.99	0.025
6.5	1.07	0.026
7.0	1.16	0.027
7.5	1.25	0.028
8.0	1.34	0.029
8.5	1.43	0.030
9.0	1.52	0.031
9.5	1.61	0.032
10.0	1.70	0.033
10.5	1.79	0.033
11.0	1.88	0.034
11.5	1.98	0.035
12.0	2.07	0.036
12.5	2.17	0.037
13.0	2.26	0.038
13.5	2.36	0.038
14.0	2.46	0.039
14.5	2.55	0.040
15.0	2.65	0.041
15.5	2.75	0.041
16.0	2.85	0.042
16.5	2.96	0.043
17.0	3.06	0.044
17.5	3.16	0.044
18.0	3.27	0.045
18.5	3.37	0.046
19.0	3.48	0.047
19.5	3.59	0.047
20.0	3.69	0.048
20.5	3.80	0.049
21.0	3.91	0.049
21.5	4.02	0.050
22.0	4.14	0.051
22.5	4.25	0.052
23.0	4.36	0.052
23.5	4.48	0.053
24.0	4.60	0.054
24.5	4.71	0.054
25.0	4.83	0.055

電流値 μA	NO ₂ 濃度 μg	NO ₂ 濃度 ppm
25.5	4.95	0.056
26.0	5.07	0.056
26.5	5.19	0.057
27.0	5.32	0.058
27.5	5.44	0.058
28.0	5.57	0.059
28.5	5.69	0.060
29.0	5.82	0.060
29.5	5.95	0.061
30.0	6.08	0.062
30.5	6.21	0.062
31.0	6.34	0.063
31.5	6.48	0.064
32.0	6.61	0.064
32.5	6.75	0.065
33.0	6.89	0.066
33.5	7.02	0.066
34.0	7.16	0.067
34.5	7.31	0.068
35.0	7.45	0.068
35.5	7.59	0.069
36.0	7.74	0.070
36.5	7.88	0.070
37.0	8.03	0.071
37.5	8.18	0.072
38.0	8.33	0.072
38.5	8.48	0.073
39.0	8.64	0.073
39.5	8.79	0.074
40.0	8.95	0.075
40.5	9.11	0.075
41.0	9.27	0.076
41.5	9.43	0.077
42.0	9.59	0.077
42.5	9.75	0.078
43.0	9.92	0.079
43.5	10.08	0.079
44.0	10.25	0.080
44.5	10.42	0.081
45.0	10.59	0.081
45.5	10.77	0.082
46.0	10.94	0.083
46.5	11.12	0.083
47.0	11.29	0.084
47.5	11.47	0.085
48.0	11.65	0.085
48.5	11.83	0.086
49.0	12.02	0.087
49.5	12.20	0.087
50.0	12.39	0.088