

八王子社  
2013年12月

検量線NO<sub>2</sub>  
9420

係数  
0.0290

定数M3  
3.1500E-05

定数M2  
3.4917E-04

定数M1  
0.17435700

電流値 μA	NO <sub>2</sub> 濃度 μg	NO <sub>2</sub> 濃度 ppm
0.0	0.00	0.000
0.5	0.09	0.009
1.0	0.17	0.012
1.5	0.26	0.015
2.0	0.35	0.017
2.5	0.44	0.019
3.0	0.53	0.021
3.5	0.62	0.023
4.0	0.71	0.024
4.5	0.79	0.026
5.0	0.88	0.027
5.5	0.97	0.029
6.0	1.07	0.030
6.5	1.16	0.031
7.0	1.25	0.032
7.5	1.34	0.034
8.0	1.43	0.035
8.5	1.53	0.036
9.0	1.62	0.037
9.5	1.71	0.038
10.0	1.81	0.039
10.5	1.91	0.040
11.0	2.00	0.041
11.5	2.10	0.042
12.0	2.20	0.043
12.5	2.30	0.044
13.0	2.39	0.045
13.5	2.49	0.046
14.0	2.60	0.047
14.5	2.70	0.048
15.0	2.80	0.049
15.5	2.90	0.049
16.0	3.01	0.050
16.5	3.11	0.051
17.0	3.22	0.052
17.5	3.33	0.053
18.0	3.44	0.054
18.5	3.54	0.055
19.0	3.65	0.055
19.5	3.77	0.056
20.0	3.88	0.057
20.5	3.99	0.058
21.0	4.11	0.059
21.5	4.22	0.060
22.0	4.34	0.060
22.5	4.46	0.061
23.0	4.58	0.062
23.5	4.70	0.063
24.0	4.82	0.064
24.5	4.94	0.064
25.0	5.07	0.065

電流値 μA	NO <sub>2</sub> 濃度 μg	NO <sub>2</sub> 濃度 ppm
		0.066
25.5	5.20	0.067
26.0	5.32	0.068
26.5	5.45	0.069
27.0	5.58	0.069
27.5	5.71	0.070
28.0	5.85	0.071
28.5	5.98	0.072
29.0	6.12	0.073
29.5	6.26	0.073
30.0	6.40	0.074
30.5	6.54	0.075
31.0	6.68	0.076
31.5	6.82	0.077
32.0	6.97	0.077
32.5	7.12	0.078
33.0	7.27	0.079
33.5	7.42	0.080
34.0	7.57	0.081
34.5	7.72	0.081
35.0	7.88	0.082
35.5	8.04	0.083
36.0	8.20	0.084
36.5	8.36	0.085
37.0	8.52	0.085
37.5	8.69	0.086
38.0	8.86	0.087
38.5	9.03	0.088
39.0	9.20	0.089
39.5	9.37	0.090
40.0	9.55	0.090
40.5	9.73	0.091
41.0	9.91	0.092
41.5	10.09	0.093
42.0	10.27	0.094
42.5	10.46	0.095
43.0	10.65	0.095
43.5	10.84	0.096
44.0	11.03	0.097
44.5	11.23	0.098
45.0	11.42	0.099
45.5	11.62	0.100
46.0	11.83	0.101
46.5	12.03	0.101
47.0	12.24	0.102
47.5	12.45	0.103
48.0	12.66	0.104
48.5	12.87	0.105
49.0	13.09	0.106
49.5	13.31	0.107
50.0	13.53	