

八王子  
2014年6月

検量線NO<sub>2</sub>  
9420

係数  
0.0190

定数M3  
1.2600E-05

定数M2  
1.1793E-03

定数M1  
0.17639377

電流値 μA	NO <sub>2</sub> 濃度 μg	NO <sub>2</sub> 濃度 ppm
0.0	0.00	0.000
0.5	0.09	0.006
1.0	0.18	0.008
1.5	0.27	0.010
2.0	0.36	0.011
2.5	0.45	0.013
3.0	0.54	0.014
3.5	0.63	0.015
4.0	0.73	0.016
4.5	0.82	0.017
5.0	0.91	0.018
5.5	1.01	0.019
6.0	1.10	0.020
6.5	1.20	0.021
7.0	1.30	0.022
7.5	1.39	0.022
8.0	1.49	0.023
8.5	1.59	0.024
9.0	1.69	0.025
9.5	1.79	0.025
10.0	1.89	0.026
10.5	2.00	0.027
11.0	2.10	0.028
11.5	2.20	0.028
12.0	2.31	0.029
12.5	2.41	0.029
13.0	2.52	0.030
13.5	2.63	0.031
14.0	2.74	0.031
14.5	2.84	0.032
15.0	2.95	0.033
15.5	3.06	0.033
16.0	3.18	0.034
16.5	3.29	0.034
17.0	3.40	0.035
17.5	3.52	0.036
18.0	3.63	0.036
18.5	3.75	0.037
19.0	3.86	0.037
19.5	3.98	0.038
20.0	4.10	0.038
20.5	4.22	0.039
21.0	4.34	0.040
21.5	4.46	0.040
22.0	4.59	0.041
22.5	4.71	0.041
23.0	4.83	0.042
23.5	4.96	0.042
24.0	5.09	0.043
24.5	5.21	0.043
25.0	5.34	0.044

電流値 μA	NO <sub>2</sub> 濃度 μg	NO <sub>2</sub> 濃度 ppm
25.5	5.47	0.044
26.0	5.60	0.045
26.5	5.74	0.046
27.0	5.87	0.046
27.5	6.00	0.047
28.0	6.14	0.047
28.5	6.28	0.048
29.0	6.41	0.048
29.5	6.55	0.049
30.0	6.69	0.049
30.5	6.83	0.050
31.0	6.98	0.050
31.5	7.12	0.051
32.0	7.27	0.051
32.5	7.41	0.052
33.0	7.56	0.052
33.5	7.71	0.053
34.0	7.86	0.053
34.5	8.01	0.054
35.0	8.16	0.054
35.5	8.31	0.055
36.0	8.47	0.055
36.5	8.62	0.056
37.0	8.78	0.056
37.5	8.94	0.057
38.0	9.10	0.057
38.5	9.26	0.058
39.0	9.42	0.058
39.5	9.58	0.059
40.0	9.75	0.059
40.5	9.92	0.060
41.0	10.08	0.060
41.5	10.25	0.061
42.0	10.42	0.061
42.5	10.59	0.062
43.0	10.77	0.062
43.5	10.94	0.063
44.0	11.12	0.063
44.5	11.30	0.064
45.0	11.47	0.064
45.5	11.65	0.065
46.0	11.84	0.065
46.5	12.02	0.066
47.0	12.20	0.066
47.5	12.39	0.067
48.0	12.58	0.067
48.5	12.77	0.068
49.0	12.96	0.068
49.5	13.15	0.069
50.0	13.34	0.069