

八王子ネット  
2015年12月

検量線NO<sub>2</sub>  
9420

係数  
0.0220

定数M3  
7.1000E-06

定数M2  
1.4222E-03

定数M1  
0.15809238

電流値 μA	NO <sub>2</sub> 濃度 μg	NO <sub>2</sub> 濃度 ppm
0.0	0.00	0.000
0.5	0.08	0.006
1.0	0.16	0.009
1.5	0.24	0.011
2.0	0.32	0.012
2.5	0.40	0.014
3.0	0.49	0.015
3.5	0.57	0.017
4.0	0.66	0.018
4.5	0.74	0.019
5.0	0.83	0.020
5.5	0.91	0.021
6.0	1.00	0.022
6.5	1.09	0.023
7.0	1.18	0.024
7.5	1.27	0.025
8.0	1.36	0.026
8.5	1.45	0.026
9.0	1.54	0.027
9.5	1.64	0.028
10.0	1.73	0.029
10.5	1.82	0.030
11.0	1.92	0.030
11.5	2.02	0.031
12.0	2.11	0.032
12.5	2.21	0.033
13.0	2.31	0.033
13.5	2.41	0.034
14.0	2.51	0.035
14.5	2.61	0.036
15.0	2.72	0.036
15.5	2.82	0.037
16.0	2.92	0.038
16.5	3.03	0.038
17.0	3.13	0.039
17.5	3.24	0.040
18.0	3.35	0.040
18.5	3.46	0.041
19.0	3.57	0.042
19.5	3.68	0.042
20.0	3.79	0.043
20.5	3.90	0.043
21.0	4.01	0.044
21.5	4.13	0.045
22.0	4.24	0.045
22.5	4.36	0.046
23.0	4.47	0.047
23.5	4.59	0.047
24.0	4.71	0.048
24.5	4.83	0.048
25.0	4.95	0.049

電流値 μA	NO <sub>2</sub> 濃度 μg	NO <sub>2</sub> 濃度 ppm
25.5	5.07	0.050
26.0	5.20	0.050
26.5	5.32	0.051
27.0	5.45	0.051
27.5	5.57	0.052
28.0	5.70	0.053
28.5	5.83	0.053
29.0	5.95	0.054
29.5	6.08	0.054
30.0	6.21	0.055
30.5	6.35	0.055
31.0	6.48	0.056
31.5	6.61	0.057
32.0	6.75	0.057
32.5	6.88	0.058
33.0	7.02	0.058
33.5	7.16	0.059
34.0	7.30	0.059
34.5	7.44	0.060
35.0	7.58	0.061
35.5	7.72	0.061
36.0	7.87	0.062
36.5	8.01	0.062
37.0	8.16	0.063
37.5	8.30	0.063
38.0	8.45	0.064
38.5	8.60	0.065
39.0	8.75	0.065
39.5	8.90	0.066
40.0	9.05	0.066
40.5	9.21	0.067
41.0	9.36	0.067
41.5	9.52	0.068
42.0	9.67	0.068
42.5	9.83	0.069
43.0	9.99	0.070
43.5	10.15	0.070
44.0	10.31	0.071
44.5	10.48	0.071
45.0	10.64	0.072
45.5	10.81	0.072
46.0	10.97	0.073
46.5	11.14	0.073
47.0	11.31	0.074
47.5	11.48	0.075
48.0	11.65	0.075
48.5	11.82	0.076
49.0	12.00	0.076
49.5	12.17	0.077
50.0	12.35	0.077