

八王子ネット  
2016年5月  
検量線NO<sub>2</sub>  
9420

係数 0.0190  
定数M3 1.6990E-05  
定数M2 5.0340E-04  
定数M1 0.19037299

電流値 μA	NO <sub>2</sub> 濃度 μg	NO <sub>2</sub> 濃度 ppm
0.0	0.00	0.000
0.5	0.10	0.006
1.0	0.19	0.008
1.5	0.29	0.010
2.0	0.38	0.012
2.5	0.48	0.013
3.0	0.58	0.014
3.5	0.67	0.016
4.0	0.77	0.017
4.5	0.87	0.018
5.0	0.97	0.019
5.5	1.07	0.020
6.0	1.16	0.020
6.5	1.26	0.021
7.0	1.36	0.022
7.5	1.46	0.023
8.0	1.56	0.024
8.5	1.66	0.024
9.0	1.77	0.025
9.5	1.87	0.026
10.0	1.97	0.027
10.5	2.07	0.027
11.0	2.18	0.028
11.5	2.28	0.029
12.0	2.39	0.029
12.5	2.49	0.030
13.0	2.60	0.031
13.5	2.70	0.031
14.0	2.81	0.032
14.5	2.92	0.032
15.0	3.03	0.033
15.5	3.13	0.034
16.0	3.24	0.034
16.5	3.35	0.035
17.0	3.47	0.035
17.5	3.58	0.036
18.0	3.69	0.036
18.5	3.80	0.037
19.0	3.92	0.038
19.5	4.03	0.038
20.0	4.14	0.039
20.5	4.26	0.039
21.0	4.38	0.040
21.5	4.49	0.040
22.0	4.61	0.041
22.5	4.73	0.041
23.0	4.85	0.042
23.5	4.97	0.042
24.0	5.09	0.043
24.5	5.22	0.043
25.0	5.34	0.044

電流値 μA	NO <sub>2</sub> 濃度 μg	NO <sub>2</sub> 濃度 ppm
25.5	5.46	0.044
26.0	5.59	0.045
26.5	5.71	0.045
27.0	5.84	0.046
27.5	5.97	0.046
28.0	6.10	0.047
28.5	6.23	0.047
29.0	6.36	0.048
29.5	6.49	0.048
30.0	6.62	0.049
30.5	6.76	0.049
31.0	6.89	0.050
31.5	7.03	0.050
32.0	7.16	0.051
32.5	7.30	0.051
33.0	7.44	0.052
33.5	7.58	0.052
34.0	7.72	0.053
34.5	7.86	0.053
35.0	8.01	0.054
35.5	8.15	0.054
36.0	8.30	0.055
36.5	8.45	0.055
37.0	8.59	0.056
37.5	8.74	0.056
38.0	8.89	0.057
38.5	9.05	0.057
39.0	9.20	0.058
39.5	9.35	0.058
40.0	9.51	0.059
40.5	9.66	0.059
41.0	9.82	0.060
41.5	9.98	0.060
42.0	10.14	0.061
42.5	10.30	0.061
43.0	10.47	0.061
43.5	10.63	0.062
44.0	10.80	0.062
44.5	10.97	0.063
45.0	11.13	0.063
45.5	11.30	0.064
46.0	11.48	0.064
46.5	11.65	0.065
47.0	11.82	0.065
47.5	12.00	0.066
48.0	12.18	0.066
48.5	12.36	0.067
49.0	12.54	0.067
49.5	12.72	0.068
50.0	12.90	0.068